



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 1 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No                      | Discipline / Group  | Quantity Measured/ Instrument | Range / Frequency | * Calibration Measurement Capability(±) | Remarks   |
|---------------------------|---|-------------------------------|-------------------|---|---|
| <b>Permanent Facility</b> |   |                               |                   |   |   |
| 1                         | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure) | A.C Current @ 50Hz            | 0.1 mA to 1 mA    | 1.0% to 0.3%                            | Using 6½ Digit Multimeter By Direct/Comparison Method |
| 2                         | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure) | A.C Current @ 50Hz            | 1 A to 10 A       | 0.30% to 0.28%                          | Using 6½ Digit Multimeter By Direct/Comparison Method |
| 3                         | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure) | A.C Current @ 50Hz            | 1 mA to 100 mA    | 0.3% to 0.19%                           | Using 6½ Digit Multimeter By Direct/Comparison Method |
| 4                         | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure) | A.C Current @ 50Hz            | 100 mA to 1 A     | 0.19% to 0.30%                          | Using 6½ Digit Multimeter By Direct/Comparison Method |
| 5                         | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure) | A.C High Voltage @ 50 Hz      | 1 kV to 5 kV      | 7.4% to 4.4%                            | Using HV Probe with DMM By Direct Method              |
| 6                         | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure) | A.C Voltage @ 50Hz            | 1 V to 1000 V     | 0.12% to 0.11%                          | Using 6½ Digit Multimeter By Direct/Comparison Method |



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 2 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No | Discipline / Group  | Quantity Measured/ Instrument   | Range / Frequency | * Calibration Measurement Capability(±) | Remarks   |
|------|---|---|-------------------|---|---|
| 7    | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure) | A.C Voltage @ 50Hz  | 10 mV to 1 V      | 0.09% to 0.12%                          | Using 6½ Digit Multimeter By Direct/Comparison Method             |
| 8    | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)  | 1 Phase/3 Phase Energy@ 50 Hz(50 to 250 V)(1A to 5A) (-0.5 to 0.5 pF) | 50 Wh to 3750 Wh  | 0.11% to 0.44%                          | Using 3 Phase Power/Energy Calibrator By Direct Method            |
| 9    | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)  | 1 Phase/3 Phase Power@ 50 Hz(50 to 250 V)(1A to 5A) (-0.5 to 0.5 pF)  | 50 W to 3750 W    | 0.34% to 0.21%                          | Using 3 Phase Power/Energy Calibrator By Direct Method            |
| 10   | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)  | A.C Current @ 50Hz  | 1 mA to 100 mA    | 0.68% to 0.61%                          | Using Multifunction Calibrator By Direct Method                   |
| 11   | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)  | A.C Current @ 50Hz  | 10 A to 800 A     | 1.55% to 1.27%                          | Using Multifunction Calibrator With Current Coil By Direct Method |
| 12   | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)  | A.C Current @ 50Hz  | 100 mA to 10 A    | 0.61% to 0.51%                          | Using Multifunction Calibrator By Direct Method                   |



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 3 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No | Discipline / Group                                       | Quantity Measured/ Instrument                | Range / Frequency | * Calibration Measurement Capability(±) | Remarks  |
|------|--|--|-------------------|---|--|
| 13   | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source) | A.C Voltage @ 50Hz                           | 10 mV to 100 mV   | 1.37 % to 0.52%                         | Using Multifunction Calibrator By Direct Method        |
| 14   | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source) | A.C Voltage @ 50Hz                           | 100 mV to 1000 V  | 0.51% to 0.51%                          | Using Multifunction Calibrator By Direct Method        |
| 15   | ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source) | Power Factor @ 50 Hz(50 to 250 V)(0.1 to 5A) | -0.5 pF to 0.5 pF | 0.012pF to 0.012pF                      | Using 3 Phase Power/Energy Calibrator By Direct Method |
| 16   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)              | D.C High Voltage                             | 1 kV to 5 kV      | 4.6% to 4.6%                            | Using HV Probe with DMM By Direct Method               |
| 17   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)              | D.C Current                                  | 0.1 mA to 1 mA    | 1.01% to 0.06%                          | Using 6½ Digit Multimeter By Direct/Comparison Method  |
| 18   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)              | D.C Current                                  | 1 A to 10 A       | 0.04% to 0.19%                          | Using 6½ Digit Multimeter By Direct/Comparison Method  |
| 19   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)              | D.C Current                                  | 1 mA to 100 mA    | 0.06% to 0.07%                          | Using 6½ Digit Multimeter By Direct/Comparison Method  |



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 4 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No | Discipline / Group                          | Quantity Measured/ Instrument | Range / Frequency | * Calibration Measurement Capability(±) | Remarks  |
|------|---|-------------------------------|-------------------|---|--|
| 20   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure) | D.C Current                   | 100 mA to 1 A     | 0.07% to 0.2%                           | Using 6½ Digit Multimeter By Direct/Comparison Method                      |
| 21   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure) | D.C Voltage                   | 1 mV to 100 mV    | 0.70% to 0.012%                         | Using 6½ Digit Multimeter By Direct/Comparison Method                      |
| 22   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure) | D.C Voltage                   | 1 V to 1000 V     | 0.20% to 0.041%                         | Using 6½ Digit Multimeter By Direct/Comparison Method                      |
| 23   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure) | D.C Voltage                   | 100 mV to 1 V     | 0.01% to 0.20%                          | Using 6½ Digit Multimeter By Direct/Comparison Method                      |
| 24   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure) | DC Resistance                 | 1 to 1            | 0.70% to 2.32%                          | Using 6½ Digit Multimeter By Direct/Comparison Method                      |
| 25   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source)  | D.C Current                   | 0.1 mA to 24 mA   | 0.80% to 0.024%                         | Using Advance Modular Calibrator/Multifunction Calibrator By Direct Method |
| 26   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source)  | D.C Current                   | 10 A to 800 A     | 1.48% to 0.51%                          | Using Multifunction Calibrator With Current Coil By Direct Method          |
| 27   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source)  | D.C Current                   | 100 mA to 10 A    | 0.61% to 0.38%                          | Using Multifunction Calibrator By Direct Method                            |



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 5 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No | Discipline / Group                         | Quantity Measured/ Instrument | Range / Frequency | * Calibration Measurement Capability( $\pm$ ) | Remarks   |
|------|--|-------------------------------|-------------------|---|---|
| 28   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Current                   | 24 mA to 100 mA   | 0.68% to 0.61%                                | Using Multifunction Calibrator By Direct Method   |
| 29   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Resistance (Discrete)     | to 100 mohm       | 0.14%   | Using Discrete Standard Resistor By Direct Method |
| 30   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Resistance (Discrete)     | to 1000 mohm      | 0.12%   | Using Discrete Standard Resistor By Direct Method |
| 31   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Resistance (Discrete)     | to 10 mohm        | 0.14%   | Using Discrete Standard Resistor By Direct Method |
| 32   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Resistance (Discrete)     | to 1 mohm         | to 0.14%                                      | Using Discrete Standard Resistor By Direct Method |
| 33   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Resistance (Discrete)     | to 10 $\mu$ ohm   | to 2.26%                                      | Using Discrete Standard Resistor By Direct Method |
| 34   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Resistance (Discrete)     | to 50 $\mu$ ohm   | to 0.60 %                                     | Using Discrete Standard Resistor By Direct Method |
| 35   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Resistance (Discrete)     | to 100 $\mu$ ohm  | to 0.55%                                      | Using Discrete Standard Resistor By Direct Method |
| 36   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source) | D.C Voltage                   | 10 mV to 100 mV   | 1.17% to 0.19%                                | Using Multifunction Calibrator By Direct Method   |



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 6 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No | Discipline / Group                                  | Quantity Measured/ Instrument | Range / Frequency  | * Calibration Measurement Capability(±) | Remarks  |
|------|---|-------------------------------|--------------------|---|--|
| 37   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source)          | D.C Voltage                   | 100 mV to 1000 V   | 0.19% to 0.12%                          | Using Multifunction Calibrator By Direct Method                        |
| 38   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source)          | DC Resistance                 | 1 to 1             | 1.40% to 2.65%                          | Using Decade Resistance Box By Direct Method                           |
| 39   | ELECTRO-TECHNICAL- DIRECT CURRENT (Source)          | Resistance                    | 1 to 100           | 2.65% to 2.84%                          | Using High Resistance Jig By Direct Method                             |
| 40   | ELECTRO-TECHNICAL- TEMPERATURE SIMULATION (Measure) | B Type thermocouple           | 600 °C to 1800 °C  | 2.47°C to 1.47°C                        | Using Advance Modular Calibrator/Universal Calibrator By Direct Method |
| 41   | ELECTRO-TECHNICAL- TEMPERATURE SIMULATION (Measure) | J Type thermocouple           | -100 °C to 1200 °C | 0.76°C to 0.80°C                        | Using Advance Modular Calibrator/Universal Calibrator By Direct Method |
| 42   | ELECTRO-TECHNICAL- TEMPERATURE SIMULATION (Measure) | K Type thermocouple           | -50 °C to 1300 °C  | 0.76°C to 0.93°C                        | Using Advance Modular Calibrator/Universal Calibrator By Direct Method |
| 43   | ELECTRO-TECHNICAL- TEMPERATURE SIMULATION (Measure) | N Type thermocouple           | -50 °C to 1300 °C  | 0.60°C to 0.60°C                        | Using Advance Modular Calibrator/Universal Calibrator By Direct Method |



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 7 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No | Discipline / Group                                 | Quantity Measured/ Instrument | Range / Frequency  | * Calibration Measurement Capability( $\pm$ ) | Remarks   |
|------|--|-------------------------------|--------------------|---|---|
| 44   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | R Type thermocouple           | 360 °C to 1700 °C  | 1.46°C to 1.47°C                              | Using Advance Modular Calibrator/Universal Calibrator By Direct Method                |
| 45   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | RTD type                      | -200 °C to 800 °C  | 0.16°C to 0.33°C                              | Using Advance Modular Calibrator/Universal Calibrator By Direct Method                |
| 46   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | S Type thermocouple           | 300 °C to 1700 °C  | 1.95°C to 1.47°C                              | Using Advance Modular Calibrator/Universal Calibrator By Direct Method                |
| 47   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | T Type thermocouple           | -50 °C to 400 °C   | 0.76°C to 0.77°C                              | Using Advance Modular Calibrator/Universal Calibrator By Direct Method                |
| 48   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)  | B Type Thermocouple           | 600 °C to 1800 °C  | 2.47°C to 2.48°C                              | Using Advance Modular Calibrator/Universal Calibrator/Process Source By Direct Method |
| 49   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)  | J Type Thermocouple           | -100 °C to 1200 °C | 0.77°C to 0.77°C                              | Using Advance Modular Calibrator/Universal Calibrator/Process Source By Direct Method |



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 8 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No | Discipline / Group                                | Quantity Measured/ Instrument | Range / Frequency | * Calibration Measurement Capability(±) | Remarks   |
|------|---|-------------------------------|-------------------|---|---|
| 50   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | K Type Thermocouple           | -50 °C to 1300 °C | 0.77°C to 0.59°C                        | Using Advance Modular Calibrator/Universal Calibrator/Process Source By Direct Method |
| 51   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | N Type Thermocouple           | -50 °C to 1300 °C | 0.60°C to 0.60°C                        | Using Advance Modular Calibrator/Universal Calibrator/Process Source By Direct Method |
| 52   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | R Type Thermocouple           | 360 °C to 1700 °C | 1.46°C to 1.46°C                        | Using Advance Modular Calibrator/Universal Calibrator/Process Source By Direct Method |
| 53   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | RTD Type                      | -200 °C to 800 °C | 0.28°C to 0.57°C                        | Using Advance Modular Calibrator/Universal Calibrator/Process Source By Direct Method |
| 54   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | S Type Thermocouple           | 300 °C to 1700 °C | 1.46°C to 1.46 °C                       | Using Advance Modular Calibrator/Universal Calibrator/Process Source By Direct Method |





# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

Laboratory Name PRISM CALIBRATION CENTRE, GF-101,F/101,101 A,B, TF-85 TO101 RUDRAKSH COMPLEX-II,, AHMEDABAD, --SELECT DISTRICT--, GUJARAT , INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2480 Page No. : 9 / 60

Validity 26/11/2019 to 25/11/2021 Last Amended on 05/12/2019

| S.No | Discipline / Group                                | Quantity Measured/ Instrument  | Range / Frequency | * Calibration Measurement Capability(±) | Remarks   |
|------|---|--|-------------------|---|---|
| 55   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | T Type Thermocouple  | -50 °C to 400 °C  | 0.76°C to 0.77°C                        | Using Advance Modular Calibrator/Universal Calibrator/Process Source By Direct Method |
| 56   | ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)     | Digital Timer,Time Totalizer,Digital Stopwatch,Totalizer,Programmable Timer.   | 1 hr to 24 hr     | 1.30Sec to 5.18Sec                      | Using Digital Time Interval Meter By Direct/Comparison Method                         |
| 57   | ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)     | Digital Timer,Time Totalizer,Digital Stopwatch,Totalizer,Programmable Timer.   | 2 mSec to 1 hr    | 0.013 Sec to 1.30Sec                    | Using Digital Time Interval Meter By Direct/Comparison Method                         |
| 58   | ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)     | Frequency  | 10 Hz to 50 kHz   | 0.04% to 0.03%                          | Using 6½ Digit Multimeter By Direct/Comparison Method                                 |
| 59   | ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)      | Frequency  | 10 Hz to 50 kHz   | 0.58% to 0.02%                          | Using Advance Modular Calibrator By Direct Method                                     |
| 60   | FLUID FLOW- FLOW MEASURING DEVICES                | All Flow Rate Metering Devices such as Digital Flow Meter, Air Flow Meter, Laminar Flow Meter/Element,Dry Gas Meter,Flow Data Logger. Mass Flow Controller, Rotameter. | 50 LPM to 300 LPM | 3.40%Rdg to 3.40%Rdg                    | Using Orifice Flow Meter By Comparison Method   |